



تاریخ: 1399/9/28 11016-945

نام آزمایشگاه: پاتوبیولوژی نور

دوره: عمومی 2-99

شماره نمونه: 2.740 991210

دانشگاه: علوم پزشکی ایران

استان: تهران

### گزارش ارزیابی کیفیت بیوشیمی

نام تست	واحد	نتیجه	میانگین گروه	CV	SDI
Glucose	mg/dL	181	187.22	4.9	-0.43
Urea	mg/dL	68	79.06	7.96	-2.45
Uric acid	mg/dL	11.1	11.12	6.82	-0.02
Creatinine	mg/dL	2.65	2.73	8.53	-0.33
Total Bilirubin	mg/dL	4.62	4.4	13.92	0.26
Direct Bilirubin	mg/dL	0.81	0.86	17.27	-0.31
Cholestrol	mg/dL	221	227.32	4.77	-0.37
Triglycerides	mg/dL	422	419.44	6.91	0.08
HDL	mg/dL	71	67.58	10.78	0.42
LDL	mg/dL	100	112.44	11.59	-0.74
Albumin	g/dL	6.5	6.08	7.38	0.91
Protein	g/dL	9.2	9.53	3.85	-0.89
ALP	IU/L	613	501.05	10.17	1.44
AST	IU/L	213	208.08	8.81	0.19
ALT	IU/L	142	157.07	8.72	-0.56
CPK	IU/L	275	295	10.04	-0.37
LDH	IU/L	680	605.19	8.49	1.55
Amylase	IU/L	386	348.92	10.25	0.92
GGT	IU/L	137	135.36	11.03	0.08
Calcium	mg/dL	12.6	12.74	2.23	-0.27
phosphorus	mg/dL	9.4	9.57	2.51	-0.23
Iron	µg/dL	183	188.42	3.9	-0.19
TIBC	µg/dL	348	388.6	17.88	-0.7
Transferrin	mg/dL	462	390.77	13.52	1.22
sodium	meq/l	161	159.63	2.95	0.17
potassium	meq/l	6	5.77	5.46	1.35
Lithium	meq/l	1.93	1.91	7.63	0.1
T4	µg/dL	7.90	8.98	9.9	-0.8
TSH	µU/mL	2.30	2.38	11.36	-0.22
Ferritin	ng/mL	54.3	51.67	12.1	0.34

SDI ≤ 0.99

1 < SDI ≤ 1.99

2 < SDI ≤ 2.99

SDI ≥ 3

عالی

قابل قبول

هشدار

غیر قابل قبول