

عنوان آزمایش : Carbamazepin

عنوان دستگاه : hplc

سری ساخت : 2-1400

عنوان کنترل :

شروع کنترل : 1400/02/29

پولد سرم


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
8.255	7.649	7.043	6.437	5.831	5.225	4.619	0.606	9.415

تاریخ	مقدار
22/10/1400	7.2
25/10/1400	7
26/10/1400	7.1
27/10/1400	7.1
28/10/1400	5.9
29/10/1400	6.6
30/10/1400	6.9

تاریخ	مقدار
12/10/1400	7.2
13/10/1400	6.7
14/10/1400	6
15/10/1400	5.9
18/10/1400	6.5
19/10/1400	5.7
20/10/1400	7.3
21/10/1400	7.1

تاریخ	مقدار
01/10/1400	8
02/10/1400	6.4
04/10/1400	6.7
05/10/1400	6
06/10/1400	6.7
07/10/1400	6.7
09/10/1400	6.5
11/10/1400	6.5

Mean: 6.683	SD: 0.549	CV%: 8.217
-------------	-----------	------------



CV% 9.415 SD 0.606 -3SD 4.619 -2SD 5.225 -1SD 5.831 Mean 6.437 1SD 7.043 2SD 7.649 3SD 8.255

چاپ

عنوان کنترل بولد سرم

اطلاعات سری ساخت : 2-1400  
 آرشیو شروع کنترل : 1400/02/29

مقدار	تاریخ
8	1400/10/01
6.4	1400/10/02
6.7	1400/10/04
6	1400/10/05
6.7	1400/10/06
6.7	1400/10/07
6.5	1400/10/09
6.5	1400/10/11
7.2	1400/10/12
6.7	1400/10/13
6	1400/10/14
5.9	1400/10/15
6.5	1400/10/18
5.7	1400/10/19
7.3	1400/10/20
7.1	1400/10/21
7.2	1400/10/22
7	1400/10/25
7.1	1400/10/26
7.1	1400/10/27
5.9	1400/10/28
6.6	1400/10/29
6.9	1400/10/30

Mean: 6.683 SD: 0.549 CV%: 8.217 6σ

