

عنوان آزمایش : Digoxin-CLIA

عنوان دستگاه : simense

سری ساخت : 40362

عنوان کنترل :

شروع کنترل : 1398/12/11

biorad II

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
2.415	2.368	2.321	2.274	2.227	2.18	2.133	0.047	2.055

تاریخ	مقدار
24/05/1400	2.29
25/05/1400	2.2
26/05/1400	2.28
30/05/1400	2.27
31/05/1400	2.31

تاریخ	مقدار
14/05/1400	2.23
16/05/1400	2.32
17/05/1400	2.25
18/05/1400	2.27
19/05/1400	2.3
20/05/1400	2.2
21/05/1400	2.23
23/05/1400	2.26

تاریخ	مقدار
02/05/1400	2.33
05/05/1400	2.33
06/05/1400	2.27
09/05/1400	2.35
10/05/1400	2.28
11/05/1400	2.3
12/05/1400	2.3
13/05/1400	2.36

Mean: 2.282	SD: 0.044	CV%: 1.944
-------------	-----------	------------



biorad II

عنوان کنترل

اطلاعات

سری ساخت : 40362

آرشیو

شروع کنترل : 1398/12/11

مقدار

تاریخ

مقدار	تاریخ
2.33	1400/05/02
2.33	1400/05/05
2.27	1400/05/06
2.35	1400/05/09
2.28	1400/05/10
2.3	1400/05/11
2.3	1400/05/12
2.36	1400/05/13
2.23	1400/05/14
2.32	1400/05/16
2.25	1400/05/17
2.27	1400/05/18
2.3	1400/05/19
2.2	1400/05/20
2.23	1400/05/21
2.26	1400/05/23
2.29	1400/05/24
2.2	1400/05/25
2.28	1400/05/26
2.27	1400/05/30
2.31	1400/05/31

Mean: 2.282

SD: 0.044

CV%: 1.944

6σ

CV% 2.055 SD 0.047 -3SD 2.133 -2SD 2.18 -1SD 2.227 Mean 2.274 1SD 2.321 2SD 2.368 3SD 2.415

چاپ

simense

