

عنوان آزمایش : Folic Acid

عنوان دستگاه : simense / سری ساخت : 40382

عنوان کنترل :

biomadII

شروع کنترل : 1399/05/28

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
11.227	10.718	10.209	9.7	9.191	8.682	8.173	0.509	5.247

تاریخ	مقدار
17/04/1400	8.84
19/04/1400	10.4
21/04/1400	8.9
22/04/1400	9.9
23/04/1400	9.5

تاریخ	مقدار
09/04/1400	9.2
10/04/1400	9.58
12/04/1400	9.3
13/04/1400	10.1
14/04/1400	9.3
15/04/1400	8.98
16/04/1400	8.81

تاریخ	مقدار
01/04/1400	10.1
02/04/1400	10
03/04/1400	10.1
05/04/1400	10.3
06/04/1400	8.9
07/04/1400	9
08/04/1400	9.78

Mean: 9.526	SD: 0.544	CV%: 5.715
-------------	-----------	------------

