

عنوان آزمایش : Folic Acid-CLIA
عنوان دستگاه : SIMENSE

سری ساخت : ۴۰۳۸۳

عنوان کنترل : biorad III
شروع کنترل : ۱۳۹۹/۱۰/۲۲

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
19.955	18.899	17.843	16.787	15.731	14.675	13.619	1.056	6.29

تاریخ	مقدار
23/12/1399	18.4
24/12/1399	18.5
25/12/1399	16.5
25/12/1399	16.5
26/12/1399	18.4
27/12/1399	15.6
28/12/1399	16.9

تاریخ	مقدار
12/12/1399	18.1
13/12/1399	15.9
14/12/1399	16.5
16/12/1399	15.7
17/12/1399	16.5
18/12/1399	16.5
19/12/1399	17.4
20/12/1399	16.8

تاریخ	مقدار
02/12/1399	15.8
03/12/1399	16.6
04/12/1399	19.1
05/12/1399	16.7
06/12/1399	16
09/12/1399	15.5
10/12/1399	16.5
11/12/1399	17.8

Mean: 16.878	SD: 1.036	CV%: 6.136
--------------	-----------	------------

