

عنوان آزمایش : Lamotrigin

عنوان دستگاه : hplc

سری ساخت : 2-1400

عنوان کنترل :

شروع کنترل : 1400/02/28

پولد سرم


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
8.283	7.681	7.079	6.477	5.875	5.273	4.671	0.602	9.297

تاریخ	مقدار
25/10/1400	6.8
26/10/1400	7
27/10/1400	6.5
28/10/1400	7.2
29/10/1400	6.8
30/10/1400	6.6


تاریخ	مقدار
13/10/1400	6.8
14/10/1400	6
15/10/1400	5.9
18/10/1400	6.4
19/10/1400	6
20/10/1400	7.5
21/10/1400	6.9
22/10/1400	7.2

تاریخ	مقدار
02/10/1400	6.7
04/10/1400	6.8
05/10/1400	6.6
06/10/1400	6.7
07/10/1400	6.5
09/10/1400	7
11/10/1400	6.6
12/10/1400	6.8

Mean: 6.695	SD: 0.392	CV%: 5.857
-------------	-----------	------------



CV% 9.297 SD 0.602 -3SD 4.671 -2SD 5.273 -1SD 5.875 Mean 6.477 1SD 7.079 2SD 7.681 3SD 8.283



عنوان کنترل بولد سرم

اطلاعات **سری ساخت : 2-1400**

آرشیو **شروع کنترل : 1400/02/28**

مقدار	تاریخ
6.7	1400/10/02
6.8	1400/10/04
6.6	1400/10/05
6.7	1400/10/06
6.5	1400/10/07
7	1400/10/09
6.6	1400/10/11
6.8	1400/10/12
6.8	1400/10/13
6	1400/10/14
5.9	1400/10/15
6.4	1400/10/18
6	1400/10/19
7.5	1400/10/20
6.9	1400/10/21
7.2	1400/10/22
6.8	1400/10/25
7	1400/10/26
6.5	1400/10/27
7.2	1400/10/28
6.8	1400/10/29
6.6	1400/10/30

Mean: 6.695 SD: 0.392 CV%: 5.857 **60**

