

عنوان آزمایش : Lamotrigin

عنوان دستگاه : hplc

سری ساخت : 2-1400

عنوان کنترل :

شروع کنترل : 1400/02/28

پولد سرم

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
7.523	7.167	6.811	6.455	6.099	5.743	5.387	0.356	5.508

تاریخ	مقدار
25/05/1400	7.2
26/05/1400	6.8
30/05/1400	7.1
31/05/1400	7.1
02/06/1400	7.1

تاریخ	مقدار
12/05/1400	6.3
13/05/1400	6.7
14/05/1400	6
16/05/1400	6.1
17/05/1400	5.4
18/05/1400	6.6
20/05/1400	7.3
23/05/1400	7

تاریخ	مقدار
02/05/1400	7.5
03/05/1400	5.7
04/05/1400	6.8
05/05/1400	6.3
06/05/1400	6.7
09/05/1400	5.9
10/05/1400	6.1
11/05/1400	6.9

Mean: 6.6	SD: 0.572	CV%: 8.664
-----------	-----------	------------



بولد سرم

عنوان کنترل

اطلاعات

سری ساخت : 2-1400

آرشیو

شروع کنترل : 1400/02/28

مقدار تاریخ

مقدار	تاریخ
7.5	1400/05/02
5.7	1400/05/03
6.8	1400/05/04
6.3	1400/05/05
6.7	1400/05/06
5.9	1400/05/09
6.1	1400/05/10
6.9	1400/05/11
6.3	1400/05/12
6.7	1400/05/13
6	1400/05/14
6.1	1400/05/16
5.4	1400/05/17
6.6	1400/05/18
7.3	1400/05/20
7	1400/05/23
7.2	1400/05/25
6.8	1400/05/26
7.1	1400/05/30
7.1	1400/05/31
7.1	1400/06/02

Mean: 6.6

SD: 0.572

CV%: 8.664

6σ

CV% 5.508 SD 0.356 -3SD 5.387 -2SD 5.743 -1SD 6.099 Mean 6.455 1SD 6.811 2SD 7.167 3SD 7.523

hplc

