

عنوان آزمایش : Phenytoin

عنوان دستگاه : hplc

سری ساخت : 2-1400

عنوان کنترل :

شروع کنترل : 1400/02/28

پولد سرم


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
4.157	3.887	3.617	3.347	3.077	2.807	2.537	0.27	8.069

تاریخ	مقدار
22/10/1400	3.6
25/10/1400	3.8
26/10/1400	3.5
27/10/1400	3.6
28/10/1400	3.5
29/10/1400	3.2
30/10/1400	3.7

تاریخ	مقدار
12/10/1400	3.5
13/10/1400	3.5
14/10/1400	3.2
15/10/1400	3.3
18/10/1400	3.4
19/10/1400	3.6
20/10/1400	3.4
21/10/1400	3.5

تاریخ	مقدار
01/10/1400	3.7
02/10/1400	2.8
04/10/1400	3.6
05/10/1400	2.7
06/10/1400	3.5
07/10/1400	3.5
09/10/1400	3.4
11/10/1400	3.4

Mean: 3.43	SD: 0.26	CV%: 7.583
------------	----------	------------





CV% 8.069 SD 0.27 -3SD 2.537 -2SD 2.807 -1SD 3.077 Mean 3.347 1SD 3.617 2SD 3.887 3SD 4.157

چاپ

عنوان کنترل بولد سرم

اطلاعات **سری ساخت : 2-1400**

آرشیو **شروع کنترل : 1400/02/28**

مقدار	تاریخ
3.7	1400/10/01
2.8	1400/10/02
3.6	1400/10/04
2.7	1400/10/05
3.5	1400/10/06
3.5	1400/10/07
3.4	1400/10/09
3.4	1400/10/11
3.5	1400/10/12
3.5	1400/10/13
3.2	1400/10/14
3.3	1400/10/15
3.4	1400/10/18
3.6	1400/10/19
3.4	1400/10/20
3.5	1400/10/21
3.6	1400/10/22
3.8	1400/10/25
3.5	1400/10/26
3.6	1400/10/27
3.5	1400/10/28
3.2	1400/10/29
3.7	1400/10/30

Mean: 3.43 SD: 0.26 CV%: 7.583 **6σ**

