

عنوان آزمایش : Troponin I-CLIA
عنوان دستگاه : abbot

100 سری ساخت : 04428

عنوان کنترل : abbot-1
شروع کنترل : 1399/05/02

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
0.023	0.022	0.021	0.02	0.019	0.018	0.017	0.001	5.12

تاریخ	مقدار
21/04/1400	0.019
22/04/1400	0.021
23/04/1400	0.02
24/04/1400	0.021
26/04/1400	0.02

تاریخ	مقدار
10/04/1400	0.02
12/04/1400	0.02
13/04/1400	0.021
14/04/1400	0.019
15/04/1400	0.018
16/04/1400	0.019
17/04/1400	0.02
19/04/1400	0.02

تاریخ	مقدار
01/04/1400	0.02
02/04/1400	0.021
03/04/1400	0.02
05/04/1400	0.021
06/04/1400	0.02
07/04/1400	0.019
08/04/1400	0.02
09/04/1400	0.018

Mean: 0.02	SD: 0.001	CV%: 4.584
------------	-----------	------------





CV% 5.12

SD 0.001

-3SD 0.017

-2SD 0.018

-1SD 0.019

Mean 0.02

1SD 0.021

2SD 0.022

3SD 0.023



abbott

