

عنوان آزمایش : Vitamin B12-CLIA
عنوان دستگاه : simense

سری ساخت : 40373

عنوان کنترل : bicrad II
شروع کنترل : 1399/05/25


3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
730.906	696.004	661.102	626.2	591.298	556.396	521.494	34.902	5.574

تاریخ	مقدار
23/04/1400	632
24/04/1400	581
26/04/1400	676
27/04/1400	689
28/04/1400	650
29/04/1400	585
31/04/1400	660

تاریخ	مقدار
12/04/1400	639
13/04/1400	630
14/04/1400	645
15/04/1400	599
16/04/1400	637
17/04/1400	614
19/04/1400	670
21/04/1400	608
22/04/1400	554


تاریخ	مقدار
01/04/1400	588
02/04/1400	660
03/04/1400	603
05/04/1400	690
06/04/1400	647
07/04/1400	649
08/04/1400	665
09/04/1400	630
10/04/1400	669

Mean: 634.8	SD: 35.691	CV%: 5.622
-------------	------------	------------





CV% 5.574 SD 34.902 -3SD 521.494 -2SD 556.396 -1SD 591.298 Mean 626.2 1SD 661.102 2SD 696.004 3SD 730.906



simense

