

عنوان آزمایش : Vitamin B12-CLIA
عنوان دستگاه : simense

سری ساخت : ۴۰۳۷۳

عنوان کنترل : biorad II
شروع کنترل : ۱۳۹۹/۰۵/۲۵

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
707.02	676.391	645.762	615.133	584.504	553.875	523.246	30.629	4.979

تاریخ	مقدار
23/12/1399	584
24/12/1399	579
25/12/1399	620
26/12/1399	635
27/12/1399	619
28/12/1399	615

تاریخ	مقدار
12/12/1399	620
13/12/1399	650
14/12/1399	645
16/12/1399	642
17/12/1399	584
18/12/1399	596
19/12/1399	619
20/12/1399	643

تاریخ	مقدار
02/12/1399	614
03/12/1399	664
04/12/1399	629
05/12/1399	589
06/12/1399	569
09/12/1399	590
10/12/1399	630
11/12/1399	632

Mean: 616.727	SD: 25.969	CV%: 4.211
---------------	------------	------------

