

عنوان آزمایش : Vitamin E

عنوان دستگاه : HPLC

سری ساخت : 0

عنوان کنترل :

شروع کنترل : 1399/05/01

پولد سرم

3SD	2SD	1SD	MEAN	-1SD	-2SD	-3SD	SD	CV%
15.354	13.667	11.98	10.293	8.606	6.919	5.232	1.687	16.386

تاریخ	مقدار
25/05/1400	10.7
26/05/1400	10.3
30/05/1400	10.6
31/05/1400	10.8

تاریخ	مقدار
13/05/1400	9.5
14/05/1400	11
17/05/1400	10.4
18/05/1400	11.5
20/05/1400	10.2
21/05/1400	10.3
24/05/1400	9.1

تاریخ	مقدار
02/05/1400	8
03/05/1400	8.5
04/05/1400	9.4
05/05/1400	8.6
09/05/1400	9
11/05/1400	10.2
12/05/1400	7.8

Mean: 9.772	SD: 1.076	CV%: 11.015
-------------	-----------	-------------

